

## Ellick Bladder Evacuator

**Description:**

Ellick bladder Evacuator-Glass

**Configuration:**

See brochure

**Intended use:**

For bladder wash during transurethral resection of prostate (TURP).

**Contraindications:**

None known.

**Caution:**

Federal (USA) law restricts this device to sale by or on the order of a physician.

**Warning / Precaution:**

- Check compatibility of adapter with endoscopes and ensure tight fittings.
- Check the bulb to fit tightly on the plastic reservoir.
- Dispose in accordance with recognized medical practice and under observance of applicable law and regulation.
- Only experienced surgeons should use this device.
- Attempt to resterilize a single use device may compromise its structural integrity and/or lead to device failure which may result in patient injury or illness.
- Storage temperature between 5°C to 30°C.

**Instruction for Use:**

1. Fill the reservoir with irrigation solution.
2. Connect the adapter to the sheath of the resectoscope, press firmly inward.
3. Fill the bladder evacuator with the appropriate volume of irrigation solution.
4. Alternatively, squeeze the bulb multiple times, to remove the tissue from the bladder. Repeat if required.

**Do not reuse, reprocess & resterilize:**

Reuse, Reprocessing & resterilization may compromise the structural integrity, can also create risk of contamination & may not give desired result or create complications, infections which may result in injury, illness or death.

**Limited Express Warranty:**

The limited express warranty as set forth herein is exclusive and in lieu of all warranties of merchantability and fitness for use, remedies, obligations and liabilities for consequential damages. The products are being sold only for the purpose described herein and such limited express warranty runs only to the original user. In no event shall ALLWIN be liable for any breach of warranty in any amount exceeding the purchase price of the product. ALLWIN reserves the right to make design changes to products without liability to incorporate said changes in ALLWIN products previously designed or sold.



Read Instruction



For Single Use Only



Do not Re-Sterilize



Use By



Sterile unless package is opened or damaged



Caution : Federal (USA) law restricts device to sale by or on the order of a physician



Sterilized using ethylene oxide



Keep out of sunlight



Keep Dry



Not made with natural rubber latex



CMC Medical Devices & Drugs S. L.  
C/Horacio Lengo N° 18, CP 29006, Malaga, Spain  
Tel.: +34 951 214 054  
E-mail : info@cmcmedicaldevices.com



Allwin Medical Devices  
3305 E. Miraloma Ave., Suite 176 Anaheim, CA 92806 (USA)  
Tel. : +1 714-572-1709 | Fax : +1 714-572-1732  
Email : info@allwinmedical.com | www.allwinmedical.com